




MODEL NO.	REF	SERIAL NO.	SN
210		17040347D	
 			
 100 - 240 VAC 50/60 Hz 0.8-0.5 A			
 ADVANCED INSTRUMENTS, INC. TWO TECHNOLOGY WAY NORWOOD, MA 02062, U.S.A. TEL: 781-320-9000	 Intertek 98214		 Electronics - Dispose Properly
 <p>CAUTION: Electrical shock hazard; do not open instrument. Refer servicing to qualified service personnel. ATTENTION: Risque de choc électrique; n'ouvrez pas l'instrument. Faire effectué l'entretien par du personnel de service qualifié. WARNING: For continued protection against fire hazard, replace fuses only with same type and rating. AVERTISSEMENT: Pour la protection contre le risque d'incendie, remplacez les fusibles par des fusibles du même type seulement.</p>	<p>WARNING: Disconnect supply before servicing. AVERTISSEMENT: Coupez l'alimentation avant l'entretien et le dépannage.</p>		
FUSE RATINGS T, 250V, 1.25A		Hot-Line® Service 1 800-225-4034 OUTSIDE U.S.A. +US 781-320-9000	
EC REP EU DIRECTIVE 98/79/EC: Winckels Medical Devices Expertise (WMDE) Bergerweg 18 6085 AT Horn The Netherlands			
MADE AND PRINTED IN U.S.A.			

Người đại diện hợp pháp của cơ sở


 GIÁM ĐỐC
Vũ Tuấn Anh