

SAO Y BẢN CHÍNH
Ngày tháng năm 20...

EU DECLARATION OF CONFORMITY

issued on the basis of the manufacturer's exclusive responsibility

Manufacturer: Karel Hrnčíř – BIOMAG ID: 18848125
Chomutice 81 VAT: CZ5702232085
507 53 Chomutice Czechia



Product: Pulsed Magnetic Therapy Device BIOMAG®
SRN: CZ-MF-000025009

Model: BIOMAG® Lumina 3D-e with applicators (trade name: Lumina 3D-e)
A6P2, A8P, A11P, A12PM, A12PL, AL21, SL29, SL30-3D, SL30-P, SL60-P, SL70

Basic UDI: BIOMAG® Lumina 3D-e – 859420828BIOMAGLUMINA3DSA
BIOMAG® Applicators – 859420828APLIKATORVN

Classification: Class IIa
Classification in accordance with Annex VIII to Rule 9 of Regulation (EU) 2017/745 of 5 April 2017, as amended.

Description: The medical device is designed for additional symptomatic treatment to support the alleviation of pain, swelling, spasms and detoxification, to improve blood circulation (vasodilation) and to accelerate healing.
It is used for various health conditions involving the musculoskeletal system, for degenerative disorders and after accidents, injuries, surgical procedures, etc.

Declaration: The Declaration of Conformity is in compliance with Regulation (EU) 2017/745, of 5 April 2017, as amended.
The manufacturer confirms the assessment of the conformity of the properties with the requirements of Regulation of the European Parliament and of the Council (EU) 2017/745, as amended, which stipulates technical requirements for medical devices, and the requirements in accordance with Act 22/1997 Coll., on technical requirements for products and on the amendment and supplementation of certain acts, as amended.
Conformity was assessed in accordance with the requirements of Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011, on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS), fulfilling the requirements of Delegated Directive (EU) 2015/863, which amends Annex II to Directive 2011/65/EU.

Evidence: Annex IX – Assessment of conformity based on a quality management system and on an assessment of the technical documentation of Regulation of the European Parliament and of the Council (EU) 2017/745, as amended.

Regulations: EN 60601-1:2006/A1:2013 EN ISO 13485:2016/AC:2018
EN 60601-1-2:2015 EN ISO 14971:2019
EN 60601-1-6:2010 EN ISO 15223-1:2021
EN 60601-1-11:2010 EN ISO 10993-1:2009/AC:2010
EN 62304:2006 EN ISO 10993-5:2009

Notification: No. 2265
3EC International a.s.
Hraničná 18
821 05 Bratislava
Slovakia

EU Certificate: No. 2024-MDR/Q5-004
dated 31.01.2024 valid until 31.01.2029



GIÁM ĐỐC
Đào Văn Bắc

Date of issue: 31.01.2024

Place of issue: Jičín



507 53 Chomutice 81
IČO 188 48 125 DIČ CZ5702232085

Revision: en REV C 24/01

Signature of responsible person:

Karel Hrnčíř, director

SAO Y BẢN CHÍNH

Ngày... tháng... năm 20.....

EU DECLARATION OF CONFORMITY

issued on the basis of the manufacturer's exclusive responsibility

Manufacturer: Karel Hrnčíř – BIOMAG ID: 18848125
Chomutice 81 VAT: CZ5702232085
507 53 Chomutice Czechia



Product: Pulsed Magnetic Therapy Device BIOMAG®
SRN: CZ-MF-000025009

Model: BIOMAG® Lumio 3D-e with applicators (trade name: Lumio 3D-e)
A1S, A3S, A6P2, A8P, A9P, A10P, A11P, A12PM, A12PL, A14P, A15P, A16P, A17P

Basic UDI: BIOMAG® Lumio 3D-e – 859420828BIOMAGLUMIO3DHU
BIOMAG® Applicators – 859420828APLIKATORVN

Classification: Class IIa
Classification in accordance with Annex VIII to Rule 9 of Regulation (EU) 2017/745 of 5 April 2017, as amended.

Description: The medical device is designed for additional symptomatic treatment to support the alleviation of pain, swelling, spasms and detoxification, to improve blood circulation (vasodilation) and to accelerate healing.
It is used for various health conditions involving the musculoskeletal system, for degenerative disorders and after accidents, injuries, surgical procedures, etc.

Declaration: The Declaration of Conformity is in compliance with Regulation (EU) 2017/745, of 5 April 2017, as amended.
The manufacturer confirms the assessment of the conformity of the properties with the requirements of Regulation of the European Parliament and of the Council (EU) 2017/745, as amended, which stipulates technical requirements for medical devices, and the requirements in accordance with Act 22/1997 Coll., on technical requirements for products and on the amendment and supplementation of certain acts, as amended.
Conformity was assessed in accordance with the requirements of Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011, on the restriction of the use of certain hazardous substances in electrical and electronic equipment (RoHS), fulfilling the requirements of Delegated Directive (EU) 2015/863, which amends Annex II to Directive 2011/65/EU.

Evidence: Annex IX – Assessment of conformity based on a quality management system and on an assessment of the technical documentation of Regulation of the European Parliament and of the Council (EU) 2017/745, as amended.

Regulations: EN 60601-1:2006/A1:2013 EN ISO 13485:2016/AC:2018
EN 60601-1-2:2015 EN ISO 14971:2019
EN 60601-1-6:2010 EN ISO 15223-1:2021
EN 60601-1-11:2010 EN ISO 10993-1:2009/AC:2010
EN 62304:2006 EN ISO 10993-5:2009

Notification: No. 2265 EU Certificate: No. 2024-MDR/QS-001
3EC International a.s. dated 31.01.2024 valid until 31.01.2029
Hraničná 18
821 05 Bratislava
Slovakia

Date of issue: 31.01.2024

Place of issue: Jičín

Revision: en REV C 24/01



© 507 53 Chomutice 81
IČO 159 49 125 DIČ CZ5702232085

Signature of responsible person: Karel Hrnčíř, director



GIÁM ĐỐC
Đào Văn Bắc

